












# Dados Biométricos

	Modelo	Injetor	Constante A Nominal	SRK T	SRK II	Barrett	Haigis	Hill-RBF	HofferQ	Holiday 1 sf
<b>iSert® 151</b>	iSert 151		118.4	118.5	118.5	1.62	A0 -0.542 A1 0.161 A2 0.204		5.30	1.52
<b>iSert® PC-60AD</b>	PC60AD		118.4	118.6	118.8	1.78	A0 -0,093 A1 -0,023 A2 0,208		5.30	1.54
<b>iSert® 255</b>	iSert 255		118.4	118.5	118.5	1.62	A0 -0.542 A1 0.161 A2 0.204		5.30	1.52
<b>Nanex™ multiSert™</b>	NY1-SP		119.2	119.1	119.1	1.94	A0 -0.2676 A1 0.2382 A2 0.1993		5.715	1.904
<b>iSert® PY-60AD</b>	PY60AD		118.4	118.6	118.6	1.67	A0 -0.093 A1 -0.023 A2 -0.208		5.30	1.54
<b>Vivinex™ Toric</b>	XY1A		118.9	119.193	119.193	1.98	A0 -1.0459 A1 0.2547 A2 0.2291	119.2	5.700	1.928
<b>Vivinex Impress™</b>	XY1EM		118.8	119.193	119.193	1.98	A0 -1.0459 A1 0.2547 A2 0.2291	119.2	5.700	1.928
<b>Gemetric™</b>	XY1-G		119.0	118.998	118.998	1.88	A0 -0.0199 A1 0.3437 A2 0.1620	119.1	5.548	1.789
<b>Gemetric™ Plus</b>	XY1-GP		119.0	118.998	118.998	1.88	A0 -0.0199 A1 0.3437 A2 0.1620	119.1	5.548	1.789
<b>Gemetric™ Toric</b>	XY1-GT		119.0	118.998	118.998	1.88	A0 -0.0199 A1 0.3437 A2 0.1620	119.1	5.548	1.789
<b>Gemetric™ Plus Toric</b>	XY1-GPT		119.0	118.998	118.998	1.88	A0 -0.0199 A1 0.3437 A2 0.1620	119.1	5.548	1.789